



HAL
open science

Assessment of thermal drift of the FLIR A325sc camera: limits and recommendations

Samy Brazane, Olivier Riou, Fabien Delaleux, Laurent Ibos, Jean-Felix
Durastanti

► To cite this version:

Samy Brazane, Olivier Riou, Fabien Delaleux, Laurent Ibos, Jean-Felix Durastanti. Assessment of thermal drift of the FLIR A325sc camera: limits and recommendations. 16th Quantitative InfraRed Thermography conference, Jul 2022, Paris, France. hal-04311848

HAL Id: hal-04311848

<https://hal.u-pec.fr/hal-04311848v1>

Submitted on 28 Nov 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Samy BRAZANE, Olivier RIOU, Fabien DELALEUX, Laurent IBOS, Jean-Felix DURASTANTI
 Centre de recherches en thermique, environnement et système (CERTES), IUT Sénart Fontainebleau, Université Paris Est Créteil - France
 Contact : samy.brazane@u-pec.fr

Abstract

Thermal drift of infrared cameras remains a major problem for the accuracy and repeatability of radiometric measurements. To control the thermal drift, we developed a cold box to adjust the temperature of the FLIR A325sc camera housing. An error diagram is proposed to estimate the optimal camera operating temperature as a function of the ambient temperature for which the thermal drift is properly managed. The effectiveness of thermal drift compensation is examined on both camera functional ranges. The use of a cold box improves the thermal drift compensation efficiency and reduces the non-repeatability of irradiance measurements.

Camera features




FLIR A325sc (data sheet)	
Detector type	Uncooled microbolometer
Detector technology	Vanadium Oxide VOx
Spectral range	7.5 - 13.0 μm
Resolution Pels	320 x 240
Detector pitch	25 μm

$$L_{\Delta\lambda}(T) = \frac{A}{\exp\left(\frac{B}{T}\right) - C}$$

Temperature range	A	B	C
OS		K ⁻¹	-
[-20°C - 120°C]	14861,6	1395,6	1,0
[0°C - 350°C]	17019,3	1436,3	1,0

Spectral Response R(λ)		[-20°C - 120°C]	[0°C - 350°C]
Central wavelength λ	[μm]	10,93	
Gaussian ratio		0,507	
Gaussian λ ₁	[μm]	8,28	
Spectral band λ ₂	[μm]	12,72	
Metrological λ ₁	[μm]	8.1(8) ± 6	8.0(4) ± 6
Spectral band λ ₂	[μm]	13.0(0) ± 1	12.8(8) ± 2



Black body

We used the Fluke 9100s thermometer calibrator. This commercial device was chosen because of its temperature range [35°C - 375°C], its inter-cavity uniformity (0.07°C @ 50°C) and its accuracy of ±0.25°C @ 50°C.



Fig. 1. Fluke 9100s Thermometer Calibrator




Fig. 2. Apparent temperature of painted cavities (black body set at 50°C)

Cavity index	# 1	# 2	# 3	# 4	# 5	# 6
Diameter Φ [mm]	1,6	3,2	4	4,8	4,8	6,4
d / (D x IFOV)	3,2	6,4	8,0	9,6	9,6	12,8
Average apparent emissivity	0,962	0,973	0,984	0,988	0,987	0,988
Reproducibility	0,004	0,005	0,007	0,005	0,006	0,006
Repeatability	0,002	0,002	0,002	0,002	0,002	0,002
Accuracy	0,004	0,005	0,004	0,004	0,004	0,004
Combined uncertainty	0,006	0,007	0,008	0,007	0,008	0,008

- Apparent emissivity of cavities is improved by coating wells using graphite based paint.
- Cavities with a size larger than 9 IFOV offer high apparent emissivity.

Thermal drift problem

In our activities, we noticed a non-repetitiveness of the temperatures detected by the thermal camera FLIR A325sc as a function of the ambient temperature.




Fig. 3. Observed deviations of $t_{BB}^{\circ} - t_{app}^{\circ}$ and t_{cam}° versus t_{amb}°

- In normal use, the camera temperature varies linearly with the ambient temperature.
- We observe an underestimation of apparent temperature about 1°C @ $t_{BB}^{\circ} = 50^{\circ}\text{C}$ when the ambient temperature varies within [10°C - 35°C].

Cold box




Fig. 4. Cold box. A thermocouple indicates the temperature of camera housing





Fig. 5. Error chart. Observed deviations $t_{BB}^{\circ} - t_{app}^{\circ}$ versus $(t_{amb}^{\circ}, t_{cam}^{\circ})$. t_{BB}° is set to 50°C


Thermal drift management

We verify the accuracy of the camera calibration within its two functional ranges : range 1: [-20; 120]°C and range 2: [0; 350]°C.



t°_amb [°C]	OS	offset [°C]	OS	mae [°C]
15°C	0,6	0,20	0,1	0,04
20°C	0,7	0,22	0,1	0,03
25°C	0,5	0,15	0,2	0,03

Table 1. Analysis of deviations within range 1



- Observed deviations within [20 – 110] °C are regular in range 1. The offset shown in table are in same order of magnitude of black body temperature uncertainty (±0.25°C). It means that the camera can be considered as calibrated over the range [20 – 110] °C. Thermal drift is well managed.

- Within range 2, deviations are consequent (2.5 OS ≡ 1°C @ 60°C) and exhibit an inflection at approximately 225°C, denoting two different behaviors (zone 1 and zone 2). 225°C correspond to a Wien wavelength (5,8 μm) for which the camera lens becomes progressively opaque.

➡ Within **zone 1**, the thermal compensation is made difficult because of the self-heating of the measurement chain due to the black body irradiance. The cold box is not able to counteract this phenomenon because it thermalizes only the housing.

➡ Within **zone 2**, compensation of thermal drift takes over. However, it is not enough to counteract the self-heating of the lens becoming more absorbent. The thermal drift is reversed with a very marked slope.

⚠ This effect can be avoided during calibration process (short exposure) but not during measuring situation for which the camera is continuously irradiated.

Conclusion

We checked the calibration of the FLIR A325sc camera. The latter is found calibrated for the range 1 [20- 110] °C. Analysis of range 2 [0-350] °C highlights two major behaviors: **1.** below 225°C, the thermal drift compensation is made difficult because of the self-heating of the measurement chain due to high intensity of black body irradiance and **2.** as the optics of the camera screens the irradiance, the compensation takes over. It is not enough to counteract the self-heating of the lens becoming more absorbent. This effect can't be avoided during measuring situation for which the camera is continuously irradiated. As these effects are systematic, corrections of object's thermosignal are possible.